

REPORT DOCUMENTATION PAGE			Form Approved OMB No. 0704-0188	
Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden, to Washington Headquarters Services, Directorate for Information Operations and Reports, 1215 Jefferson Davis Highway, Suite 1204, Arlington, VA 22202-4302, and to the Office of Management and Budget, Paperwork Reduction Project (0704-0188), Washington, DC 20503.				
1. AGENCY USE ONLY (Leave blank)	2. REPORT DATE	3. REPORT TYPE AND DATES COVERED FINAL REPORT 7/1/93-12/31/94		
4. TITLE AND SUBTITLE FEASIBILITY OF AN IN SITU FLOW CYTOMETER		5. FUNDING NUMBERS N00014-93-1-1171		
6. AUTHOR(S) Robert J. Olson and Albert J. Williams, III				
7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES) WOODS HOLE OCEANOGRAPHIC INSTITUTION WOODS HOLE, MA 02543		8. PERFORMING ORGANIZATION REPORT NUMBER W.H.O.I. Proposal No. 7800.34R		
9. SPONSORING/MONITORING AGENCY NAME(S) AND ADDRESS(ES) Scientific Officer Code 322BC ONR, 800 N. Quincy Street Arlington, VA 22217-5000		10. SPONSORING/MONITORING AGENCY REPORT NUMBER		
11. SUPPLEMENTARY NOTES				
12a. DISTRIBUTION / AVAILABILITY STATEMENT Approved for public release; distribution is unlimited		12b. DISTRIBUTION CODE		
13. ABSTRACT (Maximum 200 words) Our objective was to design an <i>in situ</i> instrument to analyze the optical properties of individual suspended particles, which are important to ocean geochemistry, optics, and biology. The approach arrived at is that of a flow cytometer modified to facilitate in situ unattended operation (use of a battery-powered solid state laser for fluorescence excitation, and of a neural network to carry out real-time data analysis). To increase reliability we use a simple ducted flow of seawater through the instrument rather than injection of seawater into the center of a particle-free fluid stream as in conventional flow cytometers. Simple ducted flow, which does not mechanically confine sample particles to a path through the central region of the focussed laser beam, requires a means of discriminating particles with acceptable trajectories from those which pass through the less-intense edges of the beam and thus give inaccurate signals. We have demonstrated that two orthogonal infrared diode laser beams can be used for this purpose: the beams can be arranged so that particles passing through both infrared beams must also pass through the center of the fluorescence excitation beam, and signals from only these particles will be collected. The instrument is now in development.				
14. SUBJECT TERMS		15. NUMBER OF PAGES		
		16. PRICE CODE		
17. SECURITY CLASSIFICATION OF REPORT U	18. SECURITY CLASSIFICATION OF THIS PAGE U	19. SECURITY CLASSIFICATION OF ABSTRACT U	20. LIMITATION OF ABSTRACT UL	

19960703 041

NSN 7540-01-280-5500

Standard Form 298 (Rev. 2-89)
Prescribed by ANSI Std. Z39-18
298-102

Final Report, ONR grant N00014-93-1-1171, Feasibility of an *in situ* flow cytometer

Robert J. Olson
Biology Department, MS 32
Woods Hole Oceanographic Institution
Woods Hole, MA 02543
phone: 508 289-2565
email: rolson@whoi.edu

Albert J. Williams, III
Applied Ocean Physics and Engineering, MS 12
Woods Hole Oceanographic Institution
Woods Hole, MA 02543
phone: 508 289-2725
email: awilliams@whoi.edu

Summary of Project

This project's objective was to evaluate designs for an *in situ* instrument to analyze the optical properties of individual suspended particles, which are important to ocean geochemistry, optics, and biology. The approach arrived at is that of a flow cytometer modified to facilitate *in situ* unattended operation (use of a battery-powered solid state laser for chlorophyll and phycoerythrin fluorescence excitation, and of a neural network to carry out real-time data analysis). To increase reliability we use a simple ducted flow of seawater through the instrument rather than injection of seawater into the center of a particle-free fluid stream as in conventional flow cytometers. Simple ducted flow, which does not mechanically confine sample particles to a path through the center of the focussed laser beam, requires a means of discriminating particles with acceptable trajectories from those which pass through the less-intense edges of the beam and thus give inaccurate signals. We have demonstrated that two orthogonal infrared diode laser beams can be used for this purpose: the beams can be arranged so that particles passing through both infrared beams must also pass through the center of the fluorescence excitation beam, and signals from only these particles will be collected. The instrument is now in development.

This project's objective was to evaluate designs and arrive at a plan for constructing an *in situ* instrument to analyze the optical properties of individual suspended particles in the ocean. The resulting design was presented as a poster in the Oceanographic Instrumentation session at the AGU Ocean Sciences meeting in February 1996, and we are now in the process of constructing an instrument based on the results of this project (with funding from DOE and NSF). We analyzed the proposed approach using a laboratory mockup system, some of whose results are shown below. We plan to use an intermediate version of the instrument in ONR's Coastal Mixing and Optics project starting this summer.

Our starting point was a flow cytometer, with which we have experience in making quantitative and rapid measurements of light scattering and fluorescence on individual phytoplankton cells. However, there are problems to be addressed before this laboratory instrument could succeed in unattended *in situ* operation. The first of these is the electrical power requirement of the laser used to excite fluorescence, which in conventional flow cytometers is typically hundreds to thousands of watts, far more than is possible to supply by batteries (during moored operation, for example). With the availability of solid state lasers (such as diode-pumped frequency-doubled Nd-YAG lasers), it is now possible to obtain adequate power for fluorescence excitation (e.g., 100 mW at 532 nm) at relatively low power cost. An instrument based on such a laser could be self-contained and battery-powered (Fig. 1).

A second problem is that for *in situ* unattended operation, an increased degree of mechanical reliability will be required. To avoid the need for injection of seawater into the center of a particle-free fluid stream as in conventional flow cytometers, we will use a simple ducted flow of seawater through the instrument. Simple ducted flow, which does not physically force sample particles through the uniform central region center of the focussed laser beam, requires a means of discriminating particles with acceptable trajectories from those which pass through the non-uniform edges of the beam and thus give inaccurate signals. We have demonstrated that two orthogonal infrared (IR) diode laser beams can be used for this purpose; with appropriate focussing, particles passing through both infrared beams will also pass through the center of the fluorescence excitation beam, and signals from only these particles will be collected (Figs. 2, 3, 4).

Laboratory tests indicated that IR diode laser beams can be focussed and aimed with sufficient precision to define an analysis region within the uniform central region of a Gaussian measuring beam. This test was first performed using a conventional flow cytometer sample injection system to provide a core stream containing fluorescent beads: it was shown that the core could be displaced far enough to eliminate IR-scattering signals without significantly affecting the signals from the main beam (i.e., the IR beam was detecting only those beads passing through the center of the main beam). In addition, when beads were added to the bulk fluid and simple ducted flow was used with a prototype detection and logic circuit, small signals (i.e., those from beads passing through the edges of the main laser beam) were selectively eliminated (Fig. 5).

A third problem, that of the large amount of data obtained from continuous multiparameter measurements of individual particles, will be addressed through use of a neural network real-time data analysis approach. By producing statistical summaries and single-parameter distributions for a limited number of particle categories, data accumulation can be kept to a reasonable level even during a long-term deployment.

The instrument is now in development.




Figure 1. Schema of *in situ* flow cytometer. The instrument's optical table is mounted to the end cap of a cylindrical housing. Seawater is pulled through the flow cell (expanded view) inside the pressure housing by a pump on the outflow end of a narrow tube. The flow cell is square, but is fused to round quartz tube on the ends to facilitate sealing by O-rings. To save space the laser beam is steered with mirrors to the center of the flow cell channel, at the focal point of the collection optics. Beam shaping optics are not shown here.




Fig. 2. Schema of the optical arrangement for the *in situ* flow cytometer, as viewed from above. A green (532-nm) beam from a frequency-doubled diode-pumped solid-state (DPSS) Nd:Yag laser is focussed on the center of the flow cell channel (180x400 μm), just as in a conventional flow cytometer. Seawater containing phytoplankton and other particles is pulled through the flow cell by a pump on the outflow side, producing laminar flow in the center of the cell.




Fig. 3. Schema of flow cell channel and laser beams, viewed from axis of main beam (green)

Two infrared (IR) diode lasers focused slightly upstream are used to detect the approach of particles that will pass through the uniform region of the green beam. One of the IR beams is nearly parallel to the main green beam, and its beam waist is focussed to a spot smaller than that of the main beam to define the central, uniform part of the main beam. The laser intensity also decreases with distance away from the beam waist on-axis; IR beam #2, oriented perpendicular to the main beam, defines an along-axis region of acceptable uniformity.




Fig. 4. Testing the optical definition of the sensing region.

The two IR lasers were aligned to the main beam while analyzing fluorescent plastic microspheres (5.2 mm diameter) using a conventional sample injection system with distilled water as sheath fluid. Signals from the three detectors (two IR forward scattering detectors and red fluorescence excited by the main beam) were displayed using a digital oscilloscope, and the positions of the beams were adjusted so that beads traversed all three beams. Then the sample pump was stopped and beads were added to the sheath reservoir, so that beads were introduced in the bulk flow. The oscilloscope was set to trigger on signals from the first IR beam, and the results of 300 events were stored. Three kinds are shown:

- (1) When both IR signals were present, fluorescence signals were also.
- (2) When the second IR signal was not present, fluorescence was usually absent as well; when a fluorescence signal was present (3), it was usually small.

For quantitative measurements of particle size and fluorescence, we should only measure signals from case (1) events.




Fig. 5. Of the beads triggering the first IR beam in Fig. 4, about 10% also triggered the second IR beam, and thus produced "acceptable" fluorescence signals. When these were compared to 100 signals triggered directly from fluorescence signals (i.e., no selection), it is obvious that the "IR-selected" signals are much larger than the unselected ones. However, the distribution of signal sizes is much broader than expected for these beads (C.V. > 15% vs 5% expected), and some small signals were acquired. These results suggest that the focus and/or position of the IR beams was not sufficiently precise to correctly define the sensing region in this first trial.